



Order form

DOC226167 D
Valid from Jan 1st 2021

DL1000 Temperature Logger

Please choose one attribute from each numbered category and add the corresponding mark to the grid on top of the table.
As a result configuration code for the Vaisala device is formed. Example of order code with typical settings:

DL1000 **V** **A** **A** **N** **N** **N** **N** **N** **N**

Selections in bold are included in the prices of the basic versions.
Selections in italic are available at an extra price.

Vaisala Data Logger Configuration		DL1000	1	2	3	4	5	6	7	8	9
1. Validation Type	<i>Validation</i>	V									
	Non Validation	S									
2. Channel 1 Function	<i>Internal Temp Probe</i>	(Standard N Range Only)	A								
Length Ext 1	<i>Temperature Stainless Steel</i>	3m (10 ft.) N Range	B								
	<i>Temperature Stainless Steel</i>	7.6 m (25 ft.) N Range	C								
	<i>Temperature Water Proof</i>	3m (10 ft.) N Range	D								
	<i>Temperature Water Proof</i>	7.6 m (25 ft.) N Range	E								
	<i>Temperature Stainless Steel</i>	3m (10 ft.) V Range	K								
	<i>Temperature Stainless Steel</i>	7.6 m (25 ft.) V Range	M								
	<i>Temperature Water Proof</i>	3m (10 ft.) V Range	P								
	<i>Temperature Water Proof</i>	7.6 m (25 ft.) V Range	R								
	Boolean Input		S								
3. Channel 2 Function	Not used		N								
Length Ext 2	<i>Temperature Stainless Steel</i>	3m (10 ft.) N Range	A								
	<i>Temperature Stainless Steel</i>	7.6 m (25 ft.) N Range	B								
	<i>Temperature Water Proof</i>	3m (10 ft.) N Range	C								
	<i>Temperature Water Proof</i>	7.6 m (25 ft.) N Range	D								
	<i>Temperature Stainless Steel</i>	3m (10 ft.) V Range	J								
	<i>Temperature Stainless Steel</i>	7.6 m (25 ft.) V Range	K								
	<i>Temperature Water Proof</i>	3m (10 ft.) V Range	M								
	<i>Temperature Water Proof</i>	7.6 m (25 ft.) V Range	P								
	Boolean Input		R								
4. Additional Calibration Points*	None		N								
	1 Additional Point**	Specify Additional Point here				1					
	2 Additional Points**	Specify Additional Points here				2					
5. Boolean Door Contact Cables	None		N								
	1 cable	7.6 m (25 ft) cable with magnetic reed relay	A								
	2 cables	7.6 m (25 ft) cable with magnetic reed relay	B								
6. Preset Calibration	Recommended	12 Months							N		
Notification Interval	Customer Specified	3 - 120 Months							1		
										N	
											N
											N

* Additional 1 week delivery

**Additional Calibration Points	Temperature Range
V Range Stainless Steel and Water Proof Probes	-90 to -40 Deg C
N Range Stainless Steel and Water Proof Probes	-25 to +70 Deg C